



Aided horizontal sound localization accuracy (SLA), quantified by an Error Index (EI), as a function of the auditory brainstem response (ABR) absolute wave-V latency. Linear regression analysis showed an increase in aided EI (i.e. worse SLA) with a longer Wave-V latency ( $r=0.92$ ,  $p=0.03$ ,  $n=5$ ).